



770-2019 ADDENDUM 3

REQUEST FOR PROPOSAL FOR SUPPLY AND INSTALLATION OF GAS CHROMATOGRAPH

URGENT

PLEASE FORWARD THIS DOCUMENT TO WHOEVER IS IN POSSESSION OF THE TENDER

ISSUED: June 8, 2020
BY: Brock Bradford
TELEPHONE NO. 204 986-4752

THIS ADDENDUM SHALL BE INCORPORATED INTO THE TENDER AND SHALL FORM A PART OF THE CONTRACT DOCUMENTS

Template Version: A20190115

Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Tender, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 10 of Form A: Bid may render your Bid non-responsive.

PART E – SPECIFICATIONS

Revise: E2.4(i) to read:

E2.4(i) have GC oven temperature control allowing fast and precise temperature ramping for optimum performance at thirty (30) degrees Celsius per minute or better up to 300°C, and twenty (20) degrees Celsius per minute or better from 300°C to 450°C;

Revise: E2.4(o) to read:

E2.4(o) have an inert flow path, split/split less capillary inlets;

Revise: E2.4(r) to read:

E2.4(r) be equipped to use hydrogen gas (H₂) as the carrier and nitrogen gas (N₂) as the makeup gas, or an approved substitute in accordance with B6.