



638-2017 ADDENDUM 1

SUPPLY AND DELIVERY OF LABORATORY GLASSWARE WASHER

URGENT

**PLEASE FORWARD THIS DOCUMENT TO
WHOEVER IS IN POSSESSION OF THE BID
OPPORTUNITY**

ISSUED: September 21, 2017
BY: Courtney Diduck
TELEPHONE NO. 204 986-4752

**THIS ADDENDUM SHALL BE INCORPORATED
INTO THE BID OPPORTUNITY AND SHALL
FORM A PART OF THE CONTRACT
DOCUMENTS**

Template Version: A20160708

Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Bid Opportunity, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 9 of Form A: Bid may render your Bid non-responsive.

PART E – SPECIFICATIONS

Revise: E2.2 (d) (v) to read: Drying; must inject produced heated force air that is at least one hundred (100) degrees Celsius in temperature or higher to dry glassware for a selected temperature and time.