



638-2017 ADDENDUM 2

SUPPLY AND DELIVERY OF LABORATORY GLASSWARE WASHER

URGENT

**PLEASE FORWARD THIS DOCUMENT TO
WHOEVER IS IN POSSESSION OF THE BID
OPPORTUNITY**

ISSUED: September 28, 2017
BY: Courtney Diduck
TELEPHONE NO. 204 986-4752

**THIS ADDENDUM SHALL BE INCORPORATED
INTO THE BID OPPORTUNITY AND SHALL
FORM A PART OF THE CONTRACT
DOCUMENTS**

Template Version: A20160708

Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Bid Opportunity, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 9 of Form A: Bid may render your Bid non-responsive.

PART D – SUPPLEMENTAL CONDITIONS

Revise: D11.1 to read:

D11.1 Goods shall be delivered within one hundred and twenty (120) Calendar Day(s) of the award of contract, f.o.b. destination, freight prepaid to:

Courtney Diduck
Analytical Branch Head
Water & Waste Department – Door 1
2230 Main Street
Winnipeg, Manitoba, R2V 4T8

Add: D11.2

D11.2 The Contractor shall confirm each delivery with the Contract Administrator or his/her designate, at least two (2) Business Days before delivery.

PART E – SPECIFICATIONS

Revise E2.1 (c) to read:

E2.1 (c) The equipment provided must be fully supported by the manufacture for parts and services from the date of purchase for ten (10) years;

Revise: E2.2 (j) to read:

E2.2 (j) Have a drop down counterbalance weight fold down door or either a motorized: fold down door when opened is horizontal or a door that moves vertically up and down;

(i) the fold down door itself should not exceed thirty-five (35) inches in length from the unit when opened horizontally;

(ii) a unit with a door that moves up and down vertically cannot exceed ninety-four (94) inches when door is open.

Questions and Answers

- Q1. Can you please confirm if full support of Parts & Services pricing for 10 years should be part of the quote?
- A1. The product is required to have parts and service available for purchase for 10 years but is not being requested at this time and is not required in the Quote.