



327-2024 ADDENDUM 2

SUPPLY AND DELIVERY OF MEDICAL GLOVES

URGENT

**PLEASE FORWARD THIS DOCUMENT TO
WHOEVER IS IN POSSESSION OF THE
BID/PROPOSAL**

ISSUED: May 6, 2024.
BY: Remilekun Obadina
TELEPHONE NO. 204 986-5125

**THIS ADDENDUM SHALL BE INCORPORATED
INTO THE BID/PROPOSAL AND SHALL FORM
A PART OF THE CONTRACT DOCUMENTS**

Template Version: Add 2024-02-01

Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Bid/Proposal, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 10 of Form A: Bid/Proposal may render your Bid/Proposal non-responsive.

PART E – SPECIFICATIONS

Add: E3.1(b) Subject to E1.3, a **Ronco Riva 8EC Nitrile Blue Exam Glove – 918L-SP50E** approved as alternative substitutes.

QUESTIONS AND ANSWERS

- Q1: Are the thicknesses of the fingertip (7.9 mil), palm (5.5 mil) and cuff (4.3 mil) absolute or required minimums?
A1: These specifications must be met.
- Q2: Would gloves that are diamond-textured on the palm and palm-side of the fingers be acceptable?
A2: No
- Q3: Would pinhole protection meeting AQL1.0 or AQL2.5 (the FDA standard) be acceptable?
A3: See above substitution.
- Q4: Will the City accept partial bids, should a proponent not have a glove size available?
A4: No partial bids will be accepted.